

### CERTIFICATE OF ANALYSIS

Page 1 of 1

#### CUSTOMER

Addmaster Ltd  
Darfin House  
Priestly Court  
Staffordshire Technology Park  
Stafford  
ST18 0AR  
UK.

#### CERTIFICATE NO.

1027740.507/9773

#### CUSTOMER REF.

2324

#### SAMPLE DETAILS

MEDISHOWER

#### DATE RECEIVED

10/02/2015

#### ORDER NO.

METHOD: Determination of Antibacterial Activity using Test Based on MOD ISO 22196

#### DATE ANALYSED

24/06/2015

#### DATE REPORTED

26/06/2015

#### RESULTS (AS CFU CM<sup>-2</sup>)

SAMPLE	SPECIES	CONTACT TIME		REDUCTION (INITIAL)	
		0 hrs	24 hrs	Log 10	%
BIOMASTER TREATED MEDISHOWER (BIOMASTER 612@2%)	<i>ESBL E coli</i>	9.8E+03	< 25.00	≥ 2.59	≥ 99.74%
POLYPROPYLENE	<i>ESBL E coli</i>	9.8E+03	1.3E+04		

Key: NS = Poor survival on control supplied.

The above data show the difference in the population following contact with the surface of the samples listed for 24 hours at 35°C under a RH of > 95% relative to the initial population.

IMSL MICROBIOLOGICAL SERVICES LTD

PALE LANE  
HARTLEY WINTNEY  
HANTS RG27 8DH  
UK

MANAGING DIRECTOR  
Peter D Askew

