

CERTIFICATE OF ANALYSIS

Page 1 of 1

CUSTOMER

Addmaster Ltd
Darfin House
Priestly Court
Staffordshire Technology Park
Stafford
ST18 0AR
UK.

CERTIFICATE NO.

1026750.257/9504

CUSTOMER REF.

2324

SAMPLE DETAILS

MEDISHOWER

DATE RECEIVED

24/02/2015

ORDER NO.

METHOD: Determination of Antibacterial Activity using Test Based on MOD ISO 22196

DATE ANALYSED 25/02/2015

DATE REPORTED

27/02/2015

RESULTS (AS CFU CM⁻²)

SAMPLE	SPECIES	CONTACT TIME		REDUCTION (CONTROL)	
		0 hrs	24 hrs	Log 10	%
POLYPROPYLENE	<i>K pneumoniae</i> CPE	1.3E+04	1.1E+05		
BIOMASTER TREATED MEDISHOWER	<i>K pneumoniae</i> CPE	1.3E+04	3.2E+03	1.5	97.17%

Key: NS = Poor survival on control supplied.

The above data show the difference in the population following contact with the surface of the samples listed for 24 hours at 35°C under a RH of > 95% relative to the control sample.

IMSL MICROBIOLOGICAL SERVICES LTD
PALE LANE
HARTLEY WINTNEY
HANTS RG27 8DH
UK

MANAGING DIRECTOR
Peter D Askew

